

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	12-05-2026 14:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	12-05-2026 14:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Labour And Employment
विभाग का नाम/Department Name	Employees State Insurance Corporation
संगठन का नाम/Organisation Name	Employees State Insurance Corporation
कार्यालय का नाम/Office Name	Esic Hospital Ezhukone Kollam
कुल मात्रा/Total Quantity	2
वस्तु श्रेणी /Item Category	ICU Ventilators (V2) (Q2)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) /Minimum Average Annual Turnover of the bidder (For 3 Years)	20 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years)	150 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	3 Year (s)
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है/MSE Relaxation for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer

बिड विवरण/Bid Details	
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेन् है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	3
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	7
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
विगत प्रदर्शन /Past Performance	30 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	50% Lowest Priced Technically Qualified Bidders
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
अनुमानित बिड मूल्य / Estimated Bid Value	4500000
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईएमडी राशि/EMD Amount	225000

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	62

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

MEDICAL SUPERINTENDENT

ESIC Hospital Ezhukone Kollam, Employees State Insurance Corporation, Employees State Insurance Corporation, Ministry of Labour and Employment
(Ms)

बोली विभाजन लागू नहीं किया गया/Bid splitting not applied.

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

4. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

5. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

6. Reverse Auction would be conducted amongst first 50% of the technically qualified bidders arranged in the order of prices from lowest to highest. Number of sellers eligible for participating in RA would be rounded off to next higher integer value if number of technically qualified bidders is odd (e.g. if 7 bids are technically qualified, then RA will be conducted amongst L-1 to L-4). In case number of technically qualified bidders are 2 or 3, RA will be between all without any elimination. If Buyer has chosen to split the bid amongst N sellers, then minimum N sellers would be taken to RA round. In case Primary products of only one OEM are left in contention for participation in RA based on lowest 50% bidders qualifying for RA, the number of sellers qualifying for RA would be increased to get at least products of one more OEM (directly participated or through its reseller) if available.

Further, if bid(s) of any seller(s) eligible for MSE preference is / are coming within price band of 15% of Non MSE L-1 or if bid of any seller(s) eligible for Make in India preference is / are coming within price band of 20% of non MII L-1, then such MSE / Make in India seller shall also be allowed to participate in the RA process.

Pre Bid Detail(s)

मूल्य भिन्नता खंड दस्तावेज़/Pre-Bid Date and Time	प्री-बिड स्थान/Pre-Bid Venue
06-05-2026 11:00:00	ESIC HOSPITAL EZHUKONE

ICU Ventilators (V2) (2 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

[* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GENERAL	Product Description	ICU Ventilator
PRODUCT INFORMATION	Patient Type	Adult & pediatric
	Type of technology	Turbine
	Tidal volume in ml	(50 ml or less) to (1500 ml or more)
	Respiration rate, breaths/min	up to 80 or more
	Trigger mechanism	Both pressure and flow
	FiO2%	21 to 100
	Inspiratory flow rate, L/min	upto 200 or more
	Inspiratory pressure, cm H2O	upto 80 or more
	IE ratio	1:10 to 4:1
	Sigh breath function	Yes
	PEEP/CPAP, cm H2O	upto 40 or more
	Pressure support, cm H2O	0 to 40 or more
	Leak compensation	Yes
	Auto 100%/Increase O2 button	Yes
	Control panel lock	Yes
Facility for double lung ventilation	Yes, No	
PATIENT ASSESSMENT TOOLS	Maximum wave forms displayed	3 (Pressure and time, volume and time , flow and time)
	Number of loops	2 loops (P-V, F-V) with facility of saving of 1 Loops for reference
	Maximum trending time in hrs	?72 hr

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Lung recruitment tools (PV loops) / OLT (Open lung tools)	Yes, No
	Lung mechanics visualization tool	Yes, No
	Capnography/CO2 monitoring	Yes, No
	Esophageal / transpulmonary pressure monitoring	Yes, No
	Modes of ventilation	Volume controlled, Pressure Controlled, Pressure Support, SIMV with Pressure support, SIMV with volume control with pressure support, CPAP/PEEP, Inverse Ratio Ventilation, Non invasive ventilation, Apnea /back up ventilation, SIMV (VC) with Pressure support, SIMV (PC) with Pressure Support, SIMV (PRVC) with Pressure Support
	APRV(Airway Pressure Release Ventilation) /Bivent/Bi level	Yes
	MMV(Mandatory Minute Ventilation) + PSV (Pressure Support Ventilation) /ASV(Adaptive Support Ventilation) or ALPV(Adaptive Lung Protection Ventilation) or Automode	Yes
	PRVC(Pressure-Regulated Volume Control ventilation)/Auto flow	Yes
	High flow oxygen therapy	Yes
	Automatic weaning system provided	Yes, No
	Volume Support	Yes, No
	Ventilation monitoring facility	The Ventilator is a Microprocessor Controlled ventilator with integrated facility for Ventilation monitoring
	Monitored / Displayed parameters	Peak inspiratory pressure, Mean airway pressure, PEEP pressure, Tidal volume, Minute volume, Spontaneous minute volume, FiO2 (analyzed %), Respiratory rate, Inspiratory time, Expiratory time, IE ratio, Plateau Pressure
	Availability of Status indicator on display	Ventilator mode, Battery life, patient data, alarm settings, clock

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Patient alarms	Low/high FiO ₂ , Low/high minute volume, Low inspiratory pressure, High pressure, Low PEEP, High PEEP, Apnea, Continuous high pressure/occlusion, Inverse IE, High respiratory rate, Breathing circuit disconnect
	Equipment alarms	Gas-supply failure, Power failure, Vent inoperative, Low battery, Self-diagnostic
	RSBI (Rapid Shallow Breathing Index)	Yes
	Expiratory Time constant (Tcexp)	Yes
	Compliance static and/or dynamic	Yes
	Inspired and/or expired resistance	Yes
	Occlusion pressure	Yes
	Inspiratory and expiratory hold	Yes
	Spontaneous frequency	Yes
	Total peep, intrinsic peep, extrinsic peep	Yes
	Expiratory block is autoclavable and no routine calibration required	Yes
	Auto compensation for ET Tube	Yes, No
	Facility for automatic compliance & Leakage compensation for circuit available	Yes
	RS 232 out put port	Yes
	USB/Memory card	Yes
	Availability of Remote alarm/display port	Yes
	Facility to Report (vent alarms and patient status)	Yes
	HL/7 Compliant out put	Yes
	Type of coloured display	LED/LCD/TFT
	Display should be touch screen	Yes
	NIV (Non Invasive Ventilation) to be possible in all modes of ventilation available	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Minimum Size of display (±2%)	12.0 (inch)
	Graphic display have automatic scaling facility for waves	Yes
HARDWARE AND ELECTRICAL REQUIREMENTS	Power supply	220-240 V , 50 Hz AC
	Provision of UPS	Yes
	Backup time in hrs	2
	Internal backup battery	Yes
	Backup time of internal battery in mins	30 mins or more
	Built in air source	Turbine (In built)
	Availability of stand alone compressor	No
	Compressor shall provide an oil free Medical air, with Peak output flow should be minimum 150 LPM and Air quality complying with ISO compressed air purity class	NA (for turbine)
	Medical Air Compressor should automatically activate in the event of wall air supply loss and replacement of internal filters should be performed without removing the compressor, to have washable air filter	NA (for turbine)
	Seller shall ensure compatibility of compressor with ventilator	NA (for Turbine)
Nebulizer with particle size less than 3 micron	Yes	
ADDITIONAL ACCESSORIES	Number of reusable silicon breathing circuits	2
	NV mask(Re usable)- Small,medium and large	2 Each
	Length of Power cord in m	? 3 m
	Flow sensor Placement	Both
	Number of flow sensor	20 Disposable
	Autoclavable exhalation valve/expiratory cassette / filter	NA for flow Sensor
	Humidifier: Servo controlled with digital monitoring of inspired gas temperature with heating wire	1

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Trolley for ventilator with circuit holding arm	Yes
CERTIFICATIONS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes
	Availability of valid Medical Device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Certification for manufacturing unit	ISO:13485 (Latest)
	Availability of Test Report for each supplied batch/product as per Medical Device Rule (MDR) 2017 as amended till date	Yes
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission or along with supplies as per buyer requirement	Yes
	Electrical Safety Standard	IEC 60601-1-2 or Equivalent BIS
	Safety and Essential Performance	IS 13450 Part 2 Section 12
MISCELLANEOUS PARAMETERS	OEM/Reseller (if supplied by reseller shall ensure uninterrupted availability of all spares for 10 years)	Yes
	Availability of toll free facility for technical support maintained by OEM or authorized agencies	Yes
	During the warranty period it shall be ensured that all the break down calls are attended with in 24 hrs and the complete details of service agents / contact details should be furnished to buyer and consignee at time of supplies	Yes
	User/Technical/Maintenance manuals to be supplied in English in hard and soft copy	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Details of equipments and procedures required for local calibration and routine maintenance to be supplied and advanced maintenance task documentation also to be furnished	Yes
	List of important spares and accessories, with their part numbers and price lists to be supplied to the buyer at the time of supplying the equipment	Yes
	Demonstration of equipment and training to be provided after completing supplies before acceptance	Yes
	The Principal Manufacturer must have direct Presence/approved service center In India	Yes
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	Yes
WARRANTY	Warranty (years)	3, 5 Or higher (year)

Additional Specification Parameters - ICU Ventilators (V2) (2 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
1.WARRANTY 2.CMC 3. DEMO	1.WARRANTY - 5 YEARS 2.CMC - 5 YEARS 3.DEMO - ONSITE PHYSICAL DEMONSTRATION OF QUOTED MODEL IS NECESSARY
4 . Compliance 5. Type of technology	4. Should comply with restrictions under rule 144(xi) of GFR 2017 5. Type of technology- TURBINE
6. ATC	6. ATC - The specifications mentioned in the ATC will supersede the specifications selected in the product bid . Golden parameters alone should not be considered for the technical evaluation . Technical evaluation is done on the basis of ATC compliance

* Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Reena G	691505,ESIC Hospital, Ezhukone, Kollam, Kerala - 691505	2	15

Buyback Detail

Item Name	ICU Ventilator
Item Description	ICU VENTILLATOR
Serial No.	ASEE-0001
Model	DRAGER SAVINA
Brand	DRAGER
Year of Manufacturing	2013
Purchase On	2013-06-08
Age	
Size	
Rating	
Quantity	1
Unit	NUMBER
Remark	NON FUNCTIONAL
Uploaded Image	View Uploaded Image

अस्वीकरण/Disclaimer

The vendor will be responsible for the un-installation of Buyback items.

Special terms and conditions-Version:1 effective from 15-04-2024 for category ICU Ventilators (V2)

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on the self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.

5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. **Generic**

Bidder financial standing: The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

3. **Generic**

Bidders shall quote only those products (Part of Service delivery) in the bid which are not obsolete in the market and has at least 7 years residual market life i.e. the offered product shall not be declared end-of-life by the OEM before this period.

4. **Generic**

Experience Criteria: The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for 3 years before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.

5. **Generic**

Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorised Reseller.

6. **Generic**

Malicious Code Certificate:

The seller should upload following certificate in the bid:-

(a) This is to certify that the Hardware and the Software being offered, as part of the contract, does not contain Embedded Malicious code that would activate procedures to :-

- (i) Inhibit the desires and designed function of the equipment.
- (ii) Cause physical damage to the user or equipment during the exploitation.
- (iii) Tap information resident or transient in the equipment/network.

(b) The firm will be considered to be in breach of the procurement contract, in case physical damage, loss of information or infringements related to copyright and Intellectual Property Right (IPRs) are caused due to activation of any such malicious code in embedded software.

7. **Generic**

Manufacturer Authorization: Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

8. **Generic**

Scope of supply includes Training: Number of employees to be trained

10
, Place for Training
Esic hospital Ezhukone

and Duration of training
2
days.

9. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

10. **Scope of Supply**

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

11. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

12. **Certificates**

ISO 9001: The bidder or the OEM of the offered products must have ISO 9001 certification.

13. **Certificates**

Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.

14. **Certificates**

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

15. **Service & Support**

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

16. **Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

17. **Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

18. **Warranty**

Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide Comprehensive Maintenance Service for next 5 years for the offered products at the rate not more than 5 % of contract price per annum. Buyer reserves the right to enter into a CMC agreement with the Successful Bidder / OEM after expiry of the Warranty period at above mentioned rate and the payment for the CMC charges would be made Biannually after rendering of the CMC Services of the relevant CMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the CMC contract when called upon by the buyer. CMC would include cost of

preventive maintenance , calibration , spare parts , break down calls etc .
(Upload the undertaking). The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 2% of total CMC value valid up to CMC period plus 2 months (if there is no other claim).

19. Warranty

Warranty period of the supplied products shall be 5 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

20. Warranty

Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

21. Warranty

Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

22. Buyback for Non PFMS / Non GPA

Buyback for Non PFMS/Non GPA:

1. GST will apply on the value of buyback and separate invoice is to be prepared by Government buyer/Consignee for buyback items offline outside the GeM system. Vendor should work out input credit for that GST applicable on the buy back while quoting the product/services.

2. Input GST Credit consideration by seller- seller shall offer the prices in the bid by the total amount of new goods including GST- total amount of old goods quoted by the firm excluding GST. This is because GST charged by buyer on buy-back value will be available for full ITC (input tax credit) on the hands of bidder. Therefore, for the purpose of quoting in the bid and evaluation, value as quoted by the bidder (excluding any GST on buy back) as indicated in column G below will be considered.

Illustrative examples for bidders are as under for understanding of the seller.

Excel sheet for explanation to Buyback ATC							
S · n o	Cost of Fresh Good (A)	%age of GST on Fresh Good as considered by seller at the time of quote (B).	GST Amount on Fresh Good (C)	Cost of Buyback item (D)	%age of GST on Buyback items (E) as fixed by buyer in the bid	GST Amount on Buyback item (F) as applicable at the time of invoice by buyer	Quote of the Bidder (G= {A+C-D})
1	100	28	28	20	8	1.6	108
2	100	9	9	20	8	1.6	89
3	110	10	11	24	8	1.92	97
Seller after the award of the GeM contract, will indicate all column A to G which will not alter their quoted price indicated in Colum-G							

3. Successful lowest bidder has to provide detailed cost break up of tendered items as well as buy back items separately after award of contract through email to the government buyer/consignee.

4. The seller of new item /service provider shall deposit the cost of invoiced buy back item with GST to the

- buyer for allowing seller to lift up the material from the buyer/consignee premise.
5. Transportation of buy back items from buyer premises shall be arranged by the seller itself without any extra cost.
 6. Government Buyer organisation /consignee will issue invoice of buy back items to the seller at the cost of buyback to be indicated to the buyer by the seller offline after receipt of order. This in no case will modify the bid in any manner.
 7. The seller will be entirely responsible for the safe disposal of old items without affecting the environment in any manner as per prevailing statutory rules and Act/ law of the land.
 8. The seller may visit the site and inspect the condition of buy back item in advance before quoting for the bid.
 9. Receipt of payment by the buyer for Old items and invoicing by buyer will be out of GeM portal.
 10. Government Buyer will upload the details of old goods along with the bid in the corrigendum with minimum price.
 11. If the minimum price (Optional) is indicated and sellers of new item who are not agreeing to take back the old item at equal to or above that minimum price should not participate in the bid.
 12. In case they participate into the bid and afterwards indicates the price of old goods, less than the minimum price shown by Government Buyer in corrigendum, EMD of the bidder will be forfeited.
 13. Item Details Consignee wise will be indicated in the corrigendum like -
 1. Consignee wise
 2. Item Description
 3. Serial Number
 4. Model
 5. Brand
 6. Year of manufacturing
 7. Purchased on
 8. Age
 9. Size, Rating, Designation
 10. Quantity
 11. Min price (Optional)
 12. Unit
 13. Working Guidelines (/Functional/Non-Functional)
 14. Remarks
 14. The successful bidder shall remove the old goods from the premise of the Government buyer/consignee at the time of the supply, installation and commisioning of fresh goods. Seller will be entitled for payment only after removal of old goods from Government buyer/consignee/user's premise and after supplying the fresh goods in terms of contract.
 15. Note: This clause of ATC shall be operated by non PFMS; non GPA buyer i.e. all PSU working through their ERP and not integrated with GPA.

23. **Buyer Added Bid Specific ATC**

Buyer Added text based ATC clauses

ADDITIONAL TERMS & CONDITIONS

A. Buy back bid

1. Item description-ICU Ventilator
2. Serial number- ASEE-0001
3. ModelDrager Savina
4. Brand-Drager
5. Purchased on-08.06.2013
6. Age -approx 13 years
- 7.Quantity -1
8. Minimum price-Rs 40000/-
- 9.Working guidelines -Non functional

S. N.	Parameter	Requirement Compliance
1	Warranty - 5years & CMC - 5 years	Compliance should be given
2	Should have ISO, BIS , CDSCO certificates and test reports	Compliance should be given
3	Quoted model should have following parameters- Lung Recruitment Tools (PV loop/OLT) Capnography /Co2 monitoring - Back up time(UPS)- 1-2 Hours	Compliance should be given
4	Additional Accessories to be provided: Reusable Silicon Breathing Circuits-10 Nos for each machine NIV Masks (Reusable)-S-3 nos ,M- 3 nos , L - 3 nos for each machine Power Cord Length \geq 3 meters for each machine Flow Sensor Placement-Distal / Proximal No.of flow sensor-Reusable-5Nos for each machine, Disposable-20 Nos for each machine Automatic Exhalation Valve/expiratory cassette/Filter-2No for each machine	Compliance should be given
5	Physical Demonstration of machine to be provided on site/ arranged by supplier anywhere within 200km from hospital, before the technical committee, failure in carrying out successful live physical demonstration and its adherence to our technical specification would lead to disqualification from bid process.	Compliance should be given
6	CMC process to be initiated by the service provider 3 months prior to the expiry of warranty period. There will be 95%uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 basis,with penalty,to extend CMC period by double the downtime period. Should provide service charges and all the spare costs including mechanical structural, electrical and electronic spares covered under under warranty and in CMC separately . These quoted rates should be valid and frozen for at least 3 years. Anything which is not separately mentioned as consumable will be considered to be a spare part and which would then be covered under warranty/CAMC.	Compliance should be given

7	Before commencement of CAMC period, the suppliers shall furnish a Performance Bank Guarantee for 2.5% of the cost of the equipment valid till 3 months extra after expiry of entire CAMC period.	Compliance should be given
8	All software updates should be provided free of cost during CAMC. In case of failure by the supplier, the Bank Guarantee of CAMC will be forfeited.	Compliance should be given
9	The payment of CAMC will be made on half yearly basis after satisfactory completion of said period duly certified by end User.	Compliance should be given
10	There should be provision for free of cost yearly calibration by the firm .There should be provision for free of cost yearly calibration by the firm .	Compliance should be given
11	Service: Should provide 24X7 on call support, There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 basis,with penalty,to extend CMC period by double the downtime period. Contact details of manufacturer, supplier and local service centre to be provided. service engineer should be stationed in South Kerala preferred, should give an affidavit regarding this with provision for penalty in case of failure..	Compliance should be given
12	Complaints should be attended properly, maximum within 12 hrs	Compliance should be given
13	Installations: Quoted make and model should be having installations at State / National / ESIC institutes.(This means National or State Institutes of excellences where the particular machine is currently functional ,certificates regarding these has to be produced)	Compliance should be given
14	Training: Should provide on-site training to our technicians on the machine and if needed regular refresher trainings on demand.	Compliance should be given
15	Should give an affidavit stating that the pricing offered to our hospital is the lowest offered by you, than to any other ESI/Govt. hospitals in India.	Compliance should be given

अस्वीकरण/Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer ,is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs

and including the modification of technical specifications and / or terms and conditions governing the bid. All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM. If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement specific to the Service, as the case may be, as provided in the Marketplace.

However, in case of Service, if any condition specified in General Terms and Conditions is contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions.](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---