

**बिड दस्तावेज़ / Bid Document**

| बिड विवरण/Bid Details  |  |
|--|--|
| बिड बंद होने की तारीख/समय /Bid End Date/Time   | 16-02-2026 11:00:00  |
| बिड खुलने की तारीख/समय /Bid Opening Date/Time  | 16-02-2026 11:30:00  |
| बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)  | 90 (Days)  |
| मंत्रालय/राज्य का नाम/Ministry/State Name  | Haryana  |
| विभाग का नाम/Department Name   | Medical Education & Research Department Haryana  |
| संगठन का नाम/Organisation Name   | Directorate Of Medical Education & Research  |
| कार्यालय का नाम/Office Name  | Nuh  |
| कुल मात्रा/Total Quantity  | 1  |
| वस्तु श्रेणी /Item Category  | Ophthalmic Ultrasound A B Scan (Q2)  |
| एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है/MSE Relaxation for Years of Experience and Turnover   | No   |
| स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover  | No   |
| विक्रेता से मांगे गए दस्तावेज़/Document required from seller   | Certificate (Requested in ATC),OEM Authorization Certificate<br>*In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid? | Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)   |
| बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension  | 1  |
| दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended  | 3  |

| बिड विवरण/Bid Details  |   |
|--|---|
| ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। /<br><b>Number of Auto Extension count</b>  | 1   |
| बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled   | Yes   |
| रिवर्स नीलामी योग्यता नियम/RA Qualification Rule   | 50% Lowest Priced Technically Qualified Bidders |
| व्यापक रखरखाव शुल्क आवश्यक / <b>Comprehensive Maintenance Charges Required</b>   | Yes   |
| बिड का प्रकार/Type of Bid  | Two Packet Bid                                  |
| तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation  | 2 Days  |
| निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM) | No  |
| मूल्यांकन पद्धति/Evaluation Method   | Total value wise evaluation                     |
| मध्यस्थता खंड/Arbitration Clause   | No  |
| सुलह खंड/Mediation Clause  | No  |

#### ईएमडी विवरण/EMD Detail

|                             |                     |
|-----------------------------|---------------------|
| एडवाइजरी बैंक/Advisory Bank | State Bank of India |
| ईएमडी राशि/EMD Amount       | 20000               |

#### ईपीबीजी विवरण /ePBG Detail

|   |                     |
|---|---------------------|
| एडवाइजरी बैंक/Advisory Bank                                       | State Bank of India |
| ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)                           | 3.00                |
| ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months). | 38                  |

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

#### लाभार्थी /Beneficiary :

Director

SHKM, Govt. Medical College, Nalhar, Nuh-Haryana (122107)  
(Dr. Mukesh Kumar)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

#### एमआईआई खरीद वरीयता / MII Purchase Preference

|   |   |
|---|---|
| एमआईआई खरीद वरीयता / MII Purchase Preference  | Yes   |
| मेक इन इंडिया विक्रेताओं को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% तक की सीमा में है / Purchase Preference to MII sellers available upto price within L1+X%  | 20  |
| मेक इन इंडिया खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MII purchase preference   | 50  |
| सार्वजनिक खरीद (मेक-इन-इंडिया को प्राथमिकता) आदेश 2017 के अनुसार केवल क्लास 1/क्लास 2 के स्थानीय आपूर्तिकर्ताओं को ही भागीदारी की अनुमति है दिनांक 16.09.2020 (समय-समय पर संशोधित एवं लागू) / Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time) | Yes, in compliance with the MII ORDER : DPIIT Order(as amended and applicable time to time) |

#### एमएसई खरीद वरीयता/MSE Purchase Preference

|  |     |
|--|-----|
| एमएसई खरीद वरीयता/MSE Purchase Preference  | Yes |
| सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% तक की सीमा में हो / Purchase Preference to MSE OEMs available upto price within L1+X% | 15  |
| सूक्ष्म और लघु उद्यम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference                          | 50  |

1. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The

buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

2. Purchase preference to Micro and Small Enterprises (MSEs) from the State of Bid Inviting Authority : Purchase preference will be given to MSEs as Micro and Small Enterprises from the State of Bid inviting Authority. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for percentage of 50 % of total quantity as defined/ decided in relevant policy.

3. Reverse Auction would be conducted amongst first 50% of the technically qualified bidders arranged in the order of prices from lowest to highest. Number of sellers eligible for participating in RA would be rounded off to next higher integer value if number of technically qualified bidders is odd (e.g. if 7 bids are technically qualified, then RA will be conducted amongst L-1 to L-4). In case number of technically qualified bidders are 2 or 3, RA will be between all without any elimination. If Buyer has chosen to split the bid amongst N sellers, then minimum N sellers would be taken to RA round. In case Primary products of only one OEM are left in contention for participation in RA based on lowest 50% bidders qualifying for RA, the number of sellers qualifying for RA would be increased to get at least products of one more OEM (directly participated or through its reseller) if available. Further, if bid(s) of any seller(s) eligible for MSE preference is / are coming within price band of 15% of Non MSE L-1 or if bid of any seller(s) eligible for Make in India preference is / are coming within price band of 20% of non MII L-1, then such MSE / Make in India seller shall also be allowed to participate in the RA process.

### Ophthalmic Ultrasound A B Scan ( 1 pieces )

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

#### तकनीकी विशिष्टियाँ /Technical Specifications

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

| विवरण/Specification | विशिष्टि का नाम /Specification Name   | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)  |
|---------------------|---------------------------------------|--|
| GENERAL FEATURES    | Product Name                          | Ophthalmic Ultrasound A-B Scan   |
| A Scan              | Probe Frequency                       | 10 MHz   |
|                     | Fixation Light                        | LED  |
|                     | Scan Mode                             | Automatic and Manual   |
|                     | Automatic Mode Range (Minimum)        | 20 - 30 mm   |
|                     | Manual Mode Range (Minimum)           | 0.5 - 30 mm  |
|                     | <b>Eye mode</b>                       | Cataract, Dense Cataract, Phakic, Aphakic, Pseudophakic  |
|                     | <b>Formulas</b>                       | Holladay, Hoffer-Q, Haigis, SRK-II, SRK-T, Post refractive correction formulae, Binhorst II                      |
|                     | <b>Measurements</b>                   | Anterior chamber depth, Vitreous length, Axial Length, Lens thickness, Individual zone velocities, Movable gates |
|                     | Measurement accuracy                  | ± 0.1 mm   |
|                     | Store - Scanning results for each eye | 5  |

| विवरण/Specification             | विशिष्टि का नाम /Specification Name                  | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)               |
|---------------------------------|--|---|
|                                 | Calibration  | Automatic with calibration cylinder   |
| B Scan                          | <b>Probe Frequency for B Scan</b>                    | 10 MHZ, 20 MHZ  |
|                                 | <b>Frames/Sec</b>                                    | 10  |
|                                 | Scan Angle   | 50 degree or more   |
|                                 | Scanning Mode  | Sector Scanning   |
|                                 | A/B Scans Capability                                 | A-Scan vector available as overlay on B-mode Images                           |
|                                 | Gray Scale   | 256   |
|                                 | Adjustable Gain                                      | 0 to 100 dB or more   |
|                                 | Adjustable TVG/TGC controls                          | Yes   |
|                                 | <b>Measurements for B Scan</b>                       | Distance, Area/ Perimeter   |
|                                 | Image post processing                                | Multiple curves processing / Pseudo-color processing curve / Flood fill color |
|                                 | Image and report exporting                           | Yes   |
|                                 | Video recordings                                     | Yes   |
| ULTRASOUND BIO MICROSCOPE (UBM) | <b>UBM probe</b>                                     | NA  |
|                                 | <b>UBM Transducer Frequency</b>                      | NA  |
|                                 | Scan angle   | NA  |
|                                 | View Zone  | NA  |
|                                 | Axial Resolutions                                    | NA  |
|                                 | Lateral Resolution                                   | NA  |
| DATA MANAGEMENT                 | Latest windows operating system                      | Yes   |
|                                 | High-resolution monitor display                      | LCD/LED/TFT   |
|                                 | Display size   | 22 inch or more   |
|                                 | <b>Touch screen display</b>                          | No  |
|                                 | LAN compatibility                                    | Yes   |
|                                 | USB ports provided                                   | 2   |
|                                 | Connectivity   | DICOM /EMR  |
|                                 | Power supply   | 200-240 VAC, 50 Hz  |
| ACCESSORIES                     | <b>Foot switch provided</b>                          | Yes   |
|                                 | Immersion shell for A scan provided                  | Yes   |
|                                 | Number of Cups for Ultrasound Biomicroscopy Provided | NA if not provided  |
|                                 | Number of Ultrasonic gel Bottle Provided             | NA if not provided  |

| विवरण/Specification | विशिष्टि का नाम /Specification Name  | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)   |
|---------------------|--|---|
|                     | Probe holder for each probe  | Yes   |
|                     | <b>Biometry tester</b>   | Yes   |
|                     | Trolley provided   | NA if not provided  |
|                     | Printer provided   | NA if not provided  |
|                     | UPS with power backup  | 30 mins   |
| CERTIFICATIONS      | Compliance to Medical Device Rules (MDR) 2017 as amended till date   | Yes   |
|                     | Availability of valid Medical Device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date | Yes   |
|                     | Certification for manufacturing unit   | ISO:13485 (Latest)  |
|                     | Availability of Test Report for each supplied batch/product as per Medical Device Rule (MDR) 2017 as amended till date   | Yes   |
|                     | Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission or along with supplies as per buyer requirement                           | Yes   |
|                     | Electrical Safety standards  | IEC-60601 or Equivalent BIS   |
|                     | WARRANTY   | <b>Warranty in years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)</b> |

| व्यापक रखरखाव / Comprehensive Maintenance          |        |
|--|--------|
| Warranty of required product                       | 3 Year |
| Comprehensive Maintenance Duration (Post Warranty) | 4 Year |

\*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

| क्र.सं./S.No. | परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer | पता/Address                             | मात्रा /Quantity | डिलीवरी के दिन/Delivery Days |
|---------------|---|---|------------------|------------------------------|
| 1             | Navneet Kumar Kaushik                                   | 122107,SHKM Govt Medical College Nalhar | 1                | 15                           |

**Special terms and conditions-Version:1 effective from 11-11-2024 for category Ophthalmic Ultrasound A B Scan**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be

marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.

10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

## क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

### 1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be  $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$ , subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

### 2. Generic

**Manufacturer Authorization:** Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

### 3. Scope of Supply

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

### 4. Certificates

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

### 5. Certificates

The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in

the bid document.

## 6. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

1. **Documents required for seller:** The bidder are advised to upload all documents mentioned in the bid documents carefully till closing date of bid.
2. **OEM Authorization Certificate:** -OEMself Certification/Authorization Certificate should be uploaded.(OEM Verified Reseller & OEM Verified Catalogue).
3. **Rate Certificate:** A rate certificate should be provided stating that the rates being charged are Special Hospital Rates and are in no way higher to those quoted/charged by your firm from the DGS&D New Delhi/Other State Govt. Departments/Health Institutions in India. In case any difference is found at a later stage your firm will refund the same accordingly.
4. **Quality Assurance:** Seller must provided test report of the product from NABL Accredited labs(if applicable).
5. **Valid License/Certificates:** The seller must uploaded the valid License/Registration certificate of sales for Medicines/Medical consumables/Medical Equipment/Medical Instrument /Software's/etc. as per Government instructions (if applicable).
6. **EMD Exemption &MSME Preference:-** EMD Exemption &MSME preference will be given to the state of bid inviting authority i.e. Haryana State only.
7. **MII Declaration :-**The seller applied for MII preference has to provide self certification for class-1 local supplier/class 2 local supplier with percentage of local content along with location(s) at which the local value addition is made.
8. **EMD:-** EMD shall be submitted in the form of DD
9. In case of re-sellers, if the reseller has not been approved as an authorized re-seller i.e. all stocks tagged as "Reseller" Not Verified by the OEM on GeM, the re-seller is required to give an undertaking as under while uploading the product, while accepting any order and while participating in any bid:

"I hereby undertake that I have made arrangements for getting the stores from authorized distributor / dealer / channel partner of the OEM of the offered product. At the time of delivery of goods, I will provide necessary chain documents to prove that the supplied goods are genuine and are being sourced from authorized distributor / dealer / channel partner of the OEM. In case of any complaint from the Buyer / Consignee about genuineness of the supplied products, I shall be responsible for providing genuine replacement supplies."
10. **Note:-**Any terms and conditions provided by GeM(By Default) which is not the requirement of the Buyer will be considered nullified.

## 7. Warranty

Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide Comprehensive Maintenance Service for next 4 years for the offered products at the rate not more than 10 % of contract price per annum. Buyer reserves the right to enter into a CMC agreement with the Successful Bidder / OEM after expiry of the Warranty period at above mentioned rate and the payment for the CMC charges would be made Biannually after rendering of the CMC Services of the relevant CMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the CMC contract when called upon by the buyer. CMC would include cost of

Hardware, Software, Electrical & Electronics Components, Software updates etc.  
(Upload the undertaking). The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 2% of total CMC value valid up to CMC period plus 2 months (if there is no other claim).

#### 8. **Warranty**

Warranty period of the supplied products shall be 3 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

#### 9. **Warranty**

Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. : The comprehensive warranty shall be covering the following scope

Hardware, Software, Electrical & Electronics Components, Software updates etc.  
(Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)

#### 10. **Warranty**

Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

#### 11. **Warranty**

Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

### **अस्वीकरण/Disclaimer**

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.

4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

## **Additional Clause For Comprehensive Maintenance Charges**

1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC.Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.

2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.

3.GST shall be included in the CMC Charges quoted.

4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula.A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.

5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.

6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available.Under this option up to 10 years can be chosen for CMC charges beyond warranty period.

7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.

8.The CMC functionality shall be available in bid only and no direct RA shall be applicable.In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field

depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

- CMC charges for 1st year after warranty period- Percentage to be indicated- A1
- CMC charges for 2nd year after warranty period- Percentage to be indicated- A2
- CMC charges for 3rd year after warranty period - Percentage to be indicated- A3
- CMC charges for 4th year after warranty period - Percentage to be indicated- A4
- CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2.The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3.In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

- (i) Number of years for which CMC required.
- (ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$C + C * \{ (A1/100)/(1.10^n) + (A2/100)/(1.10^{n+1}) + (A3/100)/(1.10^{n+2}) + (A4/100)/(1.10^{n+3}) + (A5/100)/(1.10^{n+4}) \}$  and so on

C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1,A2, A3, A4& A5 shall depend on how many years CMC selected. For 3 year CMC, only A1,A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4.CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5.The CMC charges shall be offered within range of 3 to 25% of cost of equipment.

9.Since CMC charges are to be paid only later for each year during CMC period,applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10.Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges.The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification.Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11.In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis.The equipment cost and CMC charges (year to year) shall be matched individually.

12.The CMC Contract shall be an offline contract to be handled by buyer.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13.CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

**All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.**

**For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.**

**The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:**

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**

- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

**All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.**

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---धन्यवाद/Thank You---**