

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	03-02-2026 17:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	03-02-2026 17:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Haryana
विभाग का नाम/Department Name	Health Department Haryana
संगठन का नाम/Organisation Name	N/a
कार्यालय का नाम/Office Name	University Of Health Sciences Rohtak
कुल मात्रा/Total Quantity	1
वस्तु श्रेणी /Item Category	Echo Cardiography Scanning Machine (V2) (Q2)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) /Minimum Average Annual Turnover of the bidder (For 3 Years)	40 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years)	80 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	2 Year (s)
टर्नओवर के लिए एमएसई को छूट प्राप्त है / MSE Relaxation for Turnover	Yes Complete
टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है / Startup Relaxation for Turnover	Yes Complete
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer

बिड विवरण/Bid Details	
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेन् है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	3
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	7
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
विगत प्रदर्शन /Past Performance	30 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	H1-Highest Priced Bid Elimination
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
अनुमानित बिड मूल्य /Estimated Bid Value	8000000
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईएमडी राशि/EMD Amount	160000

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	60

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई कटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

DIRECTOR, PGIMS, ROTAK

University Of Health Sciences Rohtak, Health Department Haryana, N/A,

(Director, Pgims, Rohtak)

बोली विभाजन लागू नहीं किया गया/Bid splitting not applied.

1. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover, shall upload the supporting documents to prove his eligibility for Relaxation.
2. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Bidder Turnover" as defined above subject to their meeting of quality and technical specifications. If the bidder is DPIIT Registered OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover shall upload the supporting documents to prove his eligibility for Relaxation.
3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
5. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
6. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.
7. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having

highest bid value should meet this criterion.

8. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- i. If number of technically qualified bidders are only 2 or 3.
- ii. If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- iii. In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- iv. If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1
- v. If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1

Echo Cardiography Scanning Machine (V2) (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GENERAL	Product Description	Echo Cardiography Scanning Machine
Types of Probes & Features	Echocardiography Scanning System should have an option of Cardiac (adult, pediatric), small organs, TEE, vascular	Yes
	Cardiac and Vascular Probes Configuration	TTE Phased Array, TTE Matrix Array, TEE Phased Array, TEE Matrix Array, Linear Probe
	TTE(Transthoracic) Probe Type	Adult, Pediatric, Neonatal
	Adult TTE probe Frequency in MHz	3 to 5 MHz (±2)
	Pediatric TTE probe Frequency in MHz	5 to 7.5 MHz (±2)
	Neonatal TTE probe Frequency in MHz	5 to 12MHz,(±2)
	TEE Adult Probe biplane and Multiplane frequency, in MHz	2 to 7.5 MHz (±2)
	TEE Pediatric Probe Required	Yes
	TEE Pediatric Probe frequency, in MHz	3 to 8 MHz (±2)
	Linear Probe for linear/vascular applications required	Yes
	Adult/Pediatric Linear probe frequency in MHz	5 to 10MHz (±2)
	No of Adult TTE Probes Required	1
	No of Pediatric TTE Probes Required	1

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	No of Neonatal TTE Probes Required	1
	No of Adult TEE Probes Required	1
	No of Pediatric TEE Probes Required	1
	No of Pediatric 3D/4D TEE Probes Required	NA
	No of Neonatal TEE Probes Required	NA
	No of Adult/pediatric Linear Probes supplied	NA
	Dedicated 3D/4D volume imaging for cardiology TTE ad TEE	Yes
	Live 3 D imaging in color doppler system	Yes
	Colour Doppler system with all application packages ,Quad loop for serial studies with high frame rate revies, harmonic imaging capability in all modes and integrated stress echo package digital storage and retrieval	Yes
	Number of grey shades for sharp contrast resolutions	256 or more
	High-definition acoustic zoom provision for enlarging sections of 2D and Colour flow images with more acoustic information for greater clarity and detail while maintaining an optimal frame rate	Yes
	Colour Flow Imaging Option	Yes
	Provision of Tissue Colorization (B-Colour) for improved contrast resolution	Yes
	Application software	Adult, Pediatric, Neonatal, Vascular applications and Transesophageal applications
	Cine Loop Features	Yes
	Frame grabber facility for post analysis	Yes
	Facility of Various maps for pre and post processing	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	ECG trigger facility	Yes
	User defined Preset	Yes
	Dedicated integrated dynamic stress echo package for flexible user defined protocols with stacked sub loops facility and contrast stress protocol usable for stress echocardiography	Yes
	Provision of Tissue movement colorization with quantification possibility for IHD/CAD/Heart Failure patients	Yes
	Number of transducer/probe ports	4
Scan Modes and Performance Parameters	Cardiac 3-D freehand Option	No
	Cardiac 3D/4D automatic Option	Yes
	B-mode	Yes
	M Mode	Yes
	Contrast Harmonic/Agent Imaging with angio mode	Yes
	Tissue harmonic imaging mode	Yes
	Provision of Harmonic imaging in TTE, TEE and Linear Probes	Yes
	Color Doppler imaging (CDI)	Yes
	Color Doppler 3-D/4-D option	Yes
	Power Doppler imaging (PDI)	Yes
	Power Dopple 3-D/4-D option	No
	Continuous wave	Yes
	Pulsed wave	Yes
	Duplex mode	Yes
	Triplex mode	Yes
	Tissue Doppler imaging	Yes
	Contrast-enhanced ultrasound (CEUS) option	Yes
Tissue synchronization, tissue velocity	Yes	

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Cardiac Shear wave elastography	No
	Cardiac compression elastography	No
	Gain control in two dimensions for additional level of flexibility for image quality control	Yes
	Provision for higher resolution	Real time high frequency 2D
	Probes supplied shall be in latest generation bandwidth	Yes
	Provision for higher sensitivity for low frequency doppler in all probes	Yes
	Latest generation Electronic Phased/sector array Colour Doppler system	Yes
	Number of Channels in color doppler system	5000000 or more (if channels are being designated as digital channels)
	System should be capable of generating real time live 3- D images	Yes
	Provision of Echocardiographic strain and strain-rate imaging with 2 D imaging	Yes, No
	Compactible with intracardiac Echo(ICE)	No
Image display and processing	DICOM 3 Compliant	Yes
	Adjustable transmit focus	Yes
	Automated B-mode (2-D) image	Yes
	Automated CDI image optimization	Yes
	Automated PW Doppler image optimization	Yes
	Dynamic receive focus	Yes
	Grayscale levels	256
	Image magnification (zoom)	Yes
	Real-time image	Yes
	Frozen image	Yes
	Touch Screen	Yes
	High Resolution Monitor	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Provision of Tilt and Swivel monitor should be able to view in all angles and all light conditions	Yes
	Monitor display size, cm (in)	21 Inch(±1) Or higher
	Split screen	Twin View and Live Quad Screen
	Single/dual monitors	Single with optional second console
	Depth, cm	?30
	POST PROCESSING	Yes
	SELECTABLE DYNAMIC RANGE	Yes
	SPECTRUM ANALYZER	Yes
	DIGITAL IMAGE STORAGE	Yes
	CINE LOOP PLAYBACK	Yes
	Minimum Frame Rate	>=100
	Hard Drive Storage Capacity , GB	?500 GB Or higher
	Speckle-tracking strain and strain rate	No
	Stress echo	Yes
	Exam Protocol	Yes
	Analysis and calculation packages	2-D wall motion tracking, 3-D wall motion tracking, Cardiac, Mitral valve analysis, User-programmable calculations, No
	Automated	Yes
	Digital calipers	Yes
	Exam Presets	Yes
	Customizable presets	Yes
	On-screen annotation	Yes
	Flash Memory	Yes
	USB connector	Yes
	Type of printer provided with the echo cardio machine	Thermal Printer
	Facility for high definition digital acquisition, review and editing of complete patient studies	Yes
	Connectivity to PC and Echo cardiography machine	Through IMS Software

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Provision of PC with echo cardiogram machine	Yes
	Type of processor in PC	i7 or Latest
	RAM capacity of PC in GB (hint: In caase PC not provided put zero)	>=4
	Processor speed in GHZ	>=3
	Frame Grabber incorporated	Yes
	Conectivity between Echo cardiogram and PC	Through IMS/Equivalent Software
	PC with all software inclusive interfaced with echocardiograph machine	Yes
	Inkjet printer provided with computer for reporting	Yes
	Glossy Colour Print Paper supplied	Not provided
	ECG Cable	2
	Echo cardiography sytem to be DICOM Ready	Yes
	Dicom Interfacability to HIS	Yes
	Dicom Interfacability Dicom to PACS	Yes
	Interfacability Dicom to RIS	Yes
	Echo Cardiography sytem supplied should be upgradable to next generation system on site	Yes
Power Requirments	Power input, VAC	220-240VAC, 50Hz fitted with Indian plug
	Resettable over current breaker shall be fitted for protection	Yes
	Power bakup	UPS
	Back up time in minutes	30, 60, 90, 120 Or higher (minute)
	Suitable Servo Controlled Stabilizer / CVT should be provided	Yes
CERTIFICATION AND REPORTS	Compliance to Medical Device Rules (MDR) 2017 as amended till date	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Availability of valid Medical Device license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date	Yes
	Manufacturing unit certification	ISO:13485 (Latest)
	Availability of Test Report for product as per Medical Device Rules (MDR) 2017 as amended till date	Yes
	Conformity to Performance and safety standards specific to the device	IEC 60601-2-37 or equivalent BIS
	Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission and/or along with supplies as per buyer requirement	Yes
WARRANTY	Warranty in years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	5.0 (year)
MISCELLANEOUS PARAMETERS	OEM/Reseller shall ensure uninterrupted availability of all spares for 10 years	Yes
	Availability of toll free facility for technical support maintained by OEM or authorized agencies	Yes
	User/Technical/Maintenance manuals to be supplied in English in hard and soft copy	Yes
	Details of equipments and procedures required for local calibration and routine maintenance to be supplied and advanced maintenance task documentation also to be furnished	Yes
	Supplier to perform installation, safety and operation checks before handover	Yes
	Training of users in operation and basic maintenance shall be provided	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Contact details of manufacturer, supplier and local service agent to be provided	Yes
	Certificate of calibration from the manufacturer	Yes

Additional Specification Parameters - Echo Cardiography Scanning Machine (V2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
WARRANTY AND CMC	WARRANTY MUST BE 5 YEARS AND CMC MUST BE 5 YEARS
PROBE TYPE	Adult TEE PROBE , ADULT TTE Phased Array PROBE, Pediatric TTE Phased Array PROBE
Cardiac and Vascular Probes Configuration	Adult TEE PROBE , ADULT TTE Phased Array PROBE, Pediatric TTE Phased Array PROBE
CERTIFICATION	The complete system should be USFDA OR CE European certified.

* Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Dr. Suneel Kumar	124001,PGIMS, ROHTAK 124001 University of Health Sciences, Rohtak HARYANA	1	30

Buyback Detail

Item Name	PHILIPS MODEL SONOS 7500
Item Description	PHILIPS MODEL SONOS 7500
Serial No.	US50420855
Model	MODEL SONOS 7500
Brand	PHILIPS
Year of Manufacturing	2004
Purchase On	2004-04-01
Age	
Size	1

Rating	
Quantity	1
Unit	ONE
Remark	PHILIPS MODEL SONOS 7500
Uploaded Image	View Uploaded Image
<p>अस्वीकरण/Disclaimer</p> <p>The vendor will be responsible for the un-installation of Buyback items.</p>	

Special terms and conditions-Version:1 effective from 15-04-2024 for category Echo Cardiography Scanning Machine (V2)

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or

otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.

10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/**Buyer Added Bid Specific Terms and Conditions**

1. **Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

अस्वीकरण/**Disclaimer**

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---