

बिड दस्तावेज़ / Bid Document

बिड विवरण / Bid Details	
बिड बंद होने की तारीख/समय / Bid End Date/Time	18-03-2026 14:00:00
बिड खुलने की तारीख/समय / Bid Opening Date/Time	18-03-2026 14:30:00
बिड पेशकश वैधता (बंद होने की तारीख से) / Bid Offer Validity (From End Date)	120 (Days)
मंत्रालय/राज्य का नाम / Ministry/State Name	Pmo
विभाग का नाम / Department Name	Department Of Atomic Energy
संगठन का नाम / Organisation Name	Nuclear Power Corporation Of India Limited
कार्यालय का नाम / Office Name	Kakrapar Gujarat Site
कुल मात्रा / Total Quantity	1
वस्तु श्रेणी / Item Category	Supply of Coulometer/Moisture meter with diaphragm along with accessories and Anolyte
GeMARPTS में खोजी गई स्ट्रिंग्स / Searched Strings used in GeMARPTS	Karl Fischer Titrator for oil samples
GeMARPTS में खोजा गया परिणाम / Searched Result generated in GeMARPTS	Karl Fischer Titrator (V2)
अधिसूचना के लिए चयनित प्रासंगिक श्रेणियाँ / Relevant Categories selected for notification	<ul style="list-style-type: none"> • Karl Fischer Titrator • Rapid Moisture Meter for Rapid Determination of Water Content for Soil IS - 12175 • Moisture analyser
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / MSE Relaxation for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Relaxation for Years of Experience and Turnover	No
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	3
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	7

बिड विवरण/Bid Details	
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	No
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	7 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईएमडी राशि/EMD Amount	44128

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	17

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई कटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

Site Director, Kakrapar Gujarat Site
Kakrapar Gujarat Site, Department of Atomic Energy, NUCLEAR POWER CORPORATION OF INDIA LIMITED, PMO
(Site Director)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	No
--------------------------------------------	----

एमआईआई के लिए सक्षम प्राधिकारी का विवरण:/Details of the Competent Authority for MII

सक्षम प्राधिकारी का नाम/Name of Competent Authority	SSHRI BISHWADIP DEY
सक्षम प्राधिकारी का पदनाम/Designation of Competent Authority	JOINT SECRETARY
सक्षम प्राधिकारी का कार्यालय/विभाग/प्रभाग/Office / Department / Division of Competent Authority	EXTERNAL RELATIONS & POWER (INDIAN FOREIGN SERVICES)
सीए अनुमोदन संख्या/CA Approval Number	REF NO. 407/1/2021-POWER/Vol.II/10436, Dated 08/08/2022
सक्षम प्राधिकारी अनुमोदन तिथि/Competent Authority Approval Date	08-08-2022
सक्षम प्राधिकारी द्वारा प्रदान की गई स्वीकृति का संक्षिप्त विवरण/Brief Description of the Approval Granted by Competent Authority	EXEMPTION FOR PROCUREMENT OF GOODS AND SERVICES FOR OPERATING STATIONS ITEM SR. NO. E-250 OF APENDIX-2

Competent Authority Approval for not opting Make In India Preference : [View Document](#)

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य L1+X% तक की सीमा में हो / Purchase Preference to MSE OEMs available upto price within L1+X%	15
सूक्ष्म और लघु उद्यम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference	100

1. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 100% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is

validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

एक्सेल में अपलोड किए जाने की आवश्यकता /Excel Upload Required :

GST BREAKUP - [1770966225.xlsx](#)

Supply Of Coulometer/Moisture Meter With Diaphragm Along With Accessories And Anolyte (1 pieces)

तकनीकी विशिष्टियाँ /Technical Specifications

क्रेता विशिष्टि दस्तावेज़ /Buyer Specification Document	Download
---------------------------------------------------------	--------------------------

प्रेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	प्रेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Trivedi Jayshreeben Ranchhodlal	394651,KAPS PLANT SITE, PO ANUMALA, DISTRICT TAPI, GUJARAT	1	75

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be (Increased quantity ÷ Original quantity) × Original delivery period (in days), subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. Scope of Supply

Scope of supply (Bid price to include all cost components) : Only supply of Goods

3. Certificates

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

4. Generic

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any

responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

5. Generic

Bidder shall submit the following documents along with their bid for Vendor Code Creation:

- a. Copy of PAN Card.
- b. Copy of GSTIN.
- c. Copy of Cancelled Cheque.
- d. Copy of EFT Mandate duly certified by Bank.

6. Generic

Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can match and verify the Data Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.

7. Generic

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

8. Generic

- 1. The Seller shall not assign the Contract in whole or part without obtaining the prior written consent of buyer.
- 2. The Seller shall not sub-contract the Contract in whole or part to any entity without obtaining the prior written consent of buyer.
- 3. The Seller shall, notwithstanding the consent and assignment/sub-contract, remain jointly and severally liable and responsible to buyer together with the assignee/ sub-contractor, for and in respect of the due performance of the Contract and the Sellers obligations there under.

9. Generic

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

10. Warranty

Warranty period of the supplied products shall be 1 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

11. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

SPECIFICATION FOR COULOMETER/ MOISTURE METER				
Sl. No.	Parameters	:	Value	Quantity
	COULOMETER/ MOISTURE METER			01 Nos.

1.0	Titration method	:	Coulometric Karl Fischer Titration.
2.0	Measuring range	:	10 µg to 200 mg H ₂ O
3.0	Resolution/Sensitivity	:	0.1 µg H ₂ O
4.0	Precision.		
4.1	Reproducibility At 10 µg - 1000 µg H ₂ O:	:	±3 µg (max.)
4.2	Above 1000 µg	:	0.5 % (max.)
5.0	Display of unit for moisture Display	:	ppm, µg, mg/kg, %. VGA color display, diagonal approx. 5.5" (min.)
6.0	End Point Detection	:	adjustable polarization current. Voltametric endpoint with AC.
7.0	Maximum Rotational speed of internal stirrer.	:	± (1600 - 1900) rpm.
8.0	Pump		
8.1	For addition of reagent.	:	> 600 mL/min (at 25 °C)
8.2	Aspirate	:	> 400 mL/min (at 25°C)
9.0	Power connection		
9.1	Main power connection	:	For connecting the external power supply unit.
9.2	MSB connector	:	For connecting dosing devices, stirrers or a Remote Box.
9.3	USB connector	:	Type A, for connecting USB devices.
10.0	Sensor connectors	:	One connector each for: <input type="checkbox"/> Generator electrode <input type="checkbox"/> Indicator electrode <input type="checkbox"/> Temperature sensor (Pt1000 or NTC)
11.0	Two connections nipples	:	For connecting tubing for aspirating solvent and extracting the contents of the titration cell.

12.0	Power requirement	:	100-240 VAC ($\pm 10\%$), 50-60 Hz ($\pm 3\%$) Indian plug top.
13.0	Max. current Outlet	:	1.5 A
14.0	Ambient conditions	:	+5 to +45 °C at max. 80% relative humidity, non-condensing.
15.0	Dimensions		
15.1	Width	:	220 mm (Max.)
15.2	Height without titration vessel holder	:	180 mm (Max.)
15.3	Height with titration vessel holder	:	280 mm (Max.)
16.0	Weight (including power supply unit)	:	7000 kg. (Max.)
17.0	End Point Indication	:	Visual display/acoustic beep
18.0	Titration vessel	:	Low drift cell design, no grease or PTFE sleeves should be required
19.0	Maximum titration speed	:	2.24 mg H ₂ O/minute
20.0	Start delay time	:	0-10 minutes/user selectable
21.0	End delay time	:	0-10 minutes/user selectable
22.0	Method storage	:	Minimum 8 user programmable methods can be stored.
23.0	Languages	:	English
24.0	Data Entry	:	Minimum 10 key touch pad.
25.0	Resistance to chemicals		resistant to the chemicals (no visible changes after 24 h of duration of action) like Methanol, Toluene, Ethyl acetate, Acetone, Hydrochloric acid.
26.0	Current at the electrode		auto: variable direct current 100, 200, 400 mA: pulsating direct current.
29.0	Maximum current		400 mA

30.0	Generator electrode.	Generator electrode with diaphragm 01 no.
31.0	Indicator electrode.	Indicator electrode - 01 no.
32.0	Measuring mode.	Determination with adjustable polarization current.
33.0	Temperature Measuring range	-10 - + 250 °C
34.0	Resolution	0.1 °C
35.0	Operation, dialog	Touch Control
36.0	Stirrers, titration stand	Built-in magnetic stirrer, Rod stirrer or Titration Stand
37.0	Attachment of balances, printer, PC, PC keyboard.	Via USB Slave Port,
38.0	Generation of PDF reports	Possible without PC.
39.0	Temperature sensor	Pt1000 or NTC
40.0	Method storage, sample data storage, result memory and data storage	Possible with touch pad.
41.0	Dialog languages	English should be one of the language.
42.0	Automatic titration starts	Yes
43.0	Measuring accuracyPt1000	±0.2 °C

FEATURES OF THE COULOMETER/ MOISTURE METER:

1. INTEGRATED PUMP FOR OPTIMUM CONVENIENCE

The integrated pump should enable the titration cell to be emptied and then refilled with fresh reagent at the touch of a button. **Contact with reagents is completely avoided.**

2. AUTOMATIC TITRATION START

No manual titration start should be required. The instrument should automatically begin the titration after the sample has been added. **No-touch reagent replacement.**

3. SAVE AND PRINT DATA, ETC. - WITHOUT A PC

USB interface for printer, keyboard etc. Save methods and determination data directly to a USB flash drive for uploading again at a later date.

4. PLUG-AND-PLAY OPERATION

All system components should be recognized and registered automatically upon connection.

5. REAGENT MONITORING FOR MAXIMUM SAFETY

All system components (electrodes, exchange units, etc.) and measurement results should be monitored and malfunctions area displayed - in plain text.

The reagent should be continuously monitored. Once self-defined criteria (number of added sample)

es, amount of added water, etc.) are met, the user should be prompted to replace the reagent. No touch reagent replacement.

6. KEEPING TRACK WITH KF ICONS

KF icons should show whether the instrument is still conditioning or if water content determination can now be started.

General Terms and Remarks:	
1	The supply should include complete system with all standard accessories included for starting the analysis, manuals and operable on 230 V, 50 Hz AC power supply with a 3-core cable of 2 meters length terminated in an Indian type plug top.
2	It should be equipped with an integral magnetic stirrer. External stirrer need not required to be connected for analysis.
3	Automatic analysis starts after sample injection.
4	All components of the instrument should be provided from the manufacturer and no local components should be integrated with the main instrument. Supplier should also provide the soft copy of operation and service manual.
5	The instrument, along with its all accessories, has to be supplied at Kakrapar Gujarat site.
6	As per qualification criteria only OEM of quoted item or authorized dealer/distributor/channel partner of OEM will be considered for process the tender. If the bidder is OEM he has to produce the VALID document as a proof and if the bidder is authorized dealer/distributor/channel partner of OEM, then he has to produce the VALID authorization certificate from the OEM along with the bid In absence of above mentioned documents along with the bid the bid will liable to be rejected and no correspondence will be carried out.

COMPLIANCE SHEET

SPECIFICATION FOR COULOMETER/ MOISTURE METER				
Sl. No.	Parameters	:	Value	Compliance from the bidder. (Yes/No/ Complied/Not complied/Remarks if any)
	COULOMETER/ MOISTURE METER	:	01 No. along with all the below specification and Analyte (pack of 500 ml x 6 no.) & Catholyte (pack of 5 ml X 10 Nos. total 2 packs).	
1.0	Titration method	:	Coulometric Karl Fischer Titration.	
2.0	Measuring range	:	10 µg to 200 mg H ₂ O	
3.0	Resolution/Sensitivity	:	0.1 µg H ₂ O	

4.0	Precision.		
4.1	Reproducibility At 10 µg - 1000 µg H ₂ O:	:	±3 µg (max.)
4.2	Above 1000 µg	:	0.5 % (max.)
5.0	Display of unit for moisture Display	:	ppm, µg, mg/kg, %. VGA color display, diagonal approx. 5.5" (min.)
6.0	End Point Detection	:	adjustable polarization current. Voltametric endpoint with AC.
7.0	Maximum Rotational speed of internal stirrer.		±(1600 - 1900) rpm.
8.0	Pump		
8.1	For addition of reagent.	:	> 600 mL/min (at 25 °C)
8.2	Aspirate	:	> 400 mL/min (at 25°C)
9.0	Power connection		
9.1	Main power connection	:	For connecting the external power supply unit.
9.2	MSB connector	:	For connecting dosing devices, stirrers or a Remote Box.
9.3	USB connector	:	Type A, for connecting USB devices.
10.0	Sensor connectors	:	One connector each for: <input type="checkbox"/> Generator electrode <input type="checkbox"/> Indicator electrode <input type="checkbox"/> Temperature sensor (Pt1000 or NTC)
11.0	Two connection nipples	:	For connecting tubing for aspirating solvent and extracting the contents of the titration cell.
12.0	Power requirement	:	100-240 VAC (±10%), 50-60 Hz (±3%) Indian plug top.
13.0	Max. current Outlet	:	1.5 A

14.0	Ambient conditions	:	+5 to +45 °C at max. 80% relative humidity, non-condensing.	
15.0	Dimensions			
15.1	Width	:	220 mm (Max.)	
15.2	Height without titration vessel holder	:	180 mm (Max.)	
15.3	Height with titration vessel holder	:	280 mm (Max.)	
16.0	Weight (including power supply unit)	:	7000 kg. (Max.)	
17.0	End Point Indication	:	Visual display/acoustic beep	
18.0	Titration vessel	:	Low drift cell design, no grease or PTFE sleeves should be required	
19.0	Maximum titration speed	:	2.24 mg H ₂ O/minute	
20.0	Start delay time	:	0-10 minutes/user selectable	
21.0	End delay time	:	0-10 minutes/user selectable	
22.0	Method storage	:	Minimum 8 user programmable methods can be stored.	
23.0	Languages	:	English	
24.0	Data Entry	:	Minimum 10 key touch pad.	
25.0	Resistance to chemicals		resistant to the chemicals (no visible changes after 24 h of duration of action) like Methanol, Toluene, Ethyl acetate, Acetone, Hydrochloric acid.	
26.0	Current at the electrode		auto: variable direct current 100, 200, 400 mA: pulsating direct current.	
29.0	Maximum current		400 mA	
30.0	Generator electrode.		Generator electrode with diaphragm 01 no.	

31.0	Indicator electrode.	Indicator electrode - 01 no.	
32.0	Measuring mode.	Determination with adjustable polarization current.	
33.0	Temperature Measuring range	-10 °C + 250 °C	
34.0	Resolution	0.1 °C	
35.0	Operation, dialog	Touch Control	
36.0	Stirrers, titration stand	Built-in magnetic stirrer, Rod stirrer or Titration Stand	
37.0	Attachment of balances, printer, PC, PC keyboard.	Via USB Slave Port,	
38.0	Generation of PDF reports	Possible without PC.	
39.0	Temperature sensor	Pt1000 or NTC	
40.0	Method storage, sample data storage, result memory and data storage	Possible with touch pad.	
41.0	Dialog languages	English should be one of the languages.	
42.0	Automatic titration start	Yes	
43.0	Measuring accuracy Pt1000	±0.2 °C	
	FEATURES OF THE ITEM		
1.0	The integrated pump should enable the titration cell to be emptied and then refilled with fresh reagent at the touch of a button. Contact with reagents is completely avoided.		
2.0	No manual titration start should be required. The instrument should automatically begin the titration after the sample has been added. No-touch reagent replacement.		
3.0	USB interface for printer, keyboard etc. Save methods and determination data directly to a USB flash drive for uploading again at a later date.		

4.0	All system components should be recognized and registered automatically upon connection.	
5.0	The reagent should be continuously monitored. Once self-defined criteria (number of added samples, amount of added water, etc.) are met, the user should be prompted to replace the reagent. No touch reagent replacement.	
6.0	Icons on the instrument should show whether the instrument is still conditioning or if water content determination can now be started.	
GENERAL TERMS AND REMARKS:		
1.0	The supply should include complete system with all standard accessories included for starting the analysis, manuals and operable on 230 V, 50 Hz AC power supply with a 3 core cable of 2 meters length terminated in an Indian type plug top.	
2.0	It should be equipped with an integral magnetic stirrer. External stirrer need not required to be connected for analysis.	
3.0	Automatic analysis starts after sample injection.	
4.0	All components of the instrument should be provided from the manufacturer and no local components should be integrated with the main instrument. Supplier should also provide the soft copy of operation and service manual.	
5.0	The instrument, along with its all accessories, has to be supplied at Kakarapar Gujarat site.	
6.0	As per qualification criteria only OEM of quoted item or authorized dealer/distributor/channel partner of OEM will be considered for process the tender. If the bidder is OEM he has to produce the VALID document as a proof and if the bidder is authorized dealer/distributor/channel partner of OEM, then he has to produce the VALID authorisation certificate from the OEM along with the bid In absence of above mentioned documents (to be submitted along with the bid) the bid will liable to be rejected and no correspondence will be carried out.	

ADDITIONAL TERMS & CONDITIONS

1) Unloading of the consignments reported at Consignee's location:

- i. Unloading of the consignments reported at Consignee's location is in the scope of the Buyer at Seller's risk. Hence, Sellers shall ensure to exclude all type of Unloading expenses and associated charges from the prices quoted for the items under the bid.
- ii. Seller shall ensure that the consignment/s shall be properly packed as per the Industry standards of respective categories of products to avoid damage during transportation and unloading of the consignment at Consignee's location. Seller is solely responsible for associated risk and damages, if any, that occur during the unloading of consignment at consignee's location due to improper/inadequate packing by the

Seller.

iii. Seller is responsible to provide details of Consignment shipped to Consignee in advance along with relevant documents and information of Transporter / Courier / Seller Representative (if applicable) / Driver and vehicle carrying the consignment. Sellers Representatives / Vehicle entry permit at Consignee location is subject to security clearance from Central Industrial Security Force (CISF).

iv. Buyer will put best efforts to unload the consignment with advance intimation and report preferably on or before 14:00 Hrs at NPCIL, KAPS Main Gate on any working days excluding Sundays, second Saturdays and Public holidays.

2) Bidder shall upload "GST breakup i.e., Percentage (%) of applicable GST only" and no prices shall be quoted in the breakup. The bidder shall take special care not to mix-up price details with the Part-I (Technical bid except price) and vice versa. Any violation of these conditions shall lead to rejection of the bid.

3) Qualification Requirements:

"If the bidder is OEM he has to produce the VALID document as a proof and if the bidder is authorized dealer/distributor/channel partner of OEM, then he has to produce the VALID authorization certificate from the OEM along with the bid. In absence of above mentioned documents (to be submitted along with the bid) the bidder will be liable to be rejected and no correspondence will be carried out".

4) **EMD can be submitted through SBI collect portal: following is the SBI collect navigation for payment.**

Sr. No	Particular	Details
01	Step-1	Go to SBI Collect portal and Click on PSU tab
02	Step-2	After clicking on PSU tab, Select Category: PSU-Public Sector Undertaking
03	Step-3	Type (Enter) "NUCLEAR" in search option. PSU name as Nuclear Power Corporation of India Limited and State as Gujarat will appear
04	Step-4	Fill up the requisite details and follow the procedure of bank payment.

5) Seller bank details for refund of EMD amount.

In the event, seller submitting EMD through NEFT/RTGS/DD/BANKER'S cheque mode, following information must be provided along with bid document, failing which, buyer shall not be responsible for any delay caused in returning EMD amount.

Sr. No	Particular	Details
01	Name of Account Holder	
02	Account Number	
03	Bank Name	
04	IFSC Code	
05	Bank Address	

Bank details (cancelled cheque, bank statement etc.). In support of above may please be uploaded along with bid.

Note: - Bidder shall ensure to provide details of (i) GeM Bid for EMD and (ii) GeM Contract for e-PBG transactions made online through NEFT/RTGS/SBI collect. Transactions without these details will be considered as unsolicited and NPCIL shall not be responsible for refund of credit against such transactions.

6) **MSE Purchase Preference & benefits:**

Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. **Only Manufacturers quoting for goods/products manufactured by them shall be eligible for availing benefits under the Public procurement policy for MSEs order 2012. Dealers/ distributors/sole-agents/resellers/Traders/Stockists will not be considered for benefits under the subject policy and are required to pay requisite EMD in case stipulated in the bid document. Seller declaration stating that the offered product is manufactured by them shall be submitted along with the bid.**

As mentioned in Section 7(4) of Ministry of MSME's Notification No. S.O2119(E) dated 26th June, 2020, an enterprise registered with any other organization under the Ministry of MSME shall register itself under **Udyam Registration**. With effect from 01.07.2020, MSEs registered under **Udyam Registration** are eligible to avail the benefits under the Policy.

All the MSEs interested in availing benefits must upload the following, failing which, such MSEs **shall not be entitled to claim the benefits** under the MSME policy:

- a. **Udyam Registration Certification (URC)**
- b. **Seller declaration that the offered product is manufactured by them**

7) **Civil Liability for Nuclear Damages (CLND) Act 2010 & Rule 2011 thereof:**

Subsequent to the enactment of CLND Act 2010 and Rule 2011, the Purchaser shall have Right to Recourse against the contractor in accordance with provisions under Section 17(a) of Civil Liability for Nuclear Damage Act, 2010.

Right of Recourse under Civil Liability for Nuclear Damages Act 2010 & Rule 2011 thereof;

- a. The Purchaser shall have Right to Recourse against the supplier in accordance with provisions under Section 17(a) of Civil Liability for Nuclear Damage Act, 2010, with following limitations, as stipulated in Rule 24 of the Civil Liability for Nuclear Damage Rules, 2011:

The Supplier's liability shall be to the extent of the Operator's liability under sub-section(2) of Section 6 of the Act or the value of the contract, whichever is less,

AND

The duration of Supplier's liability shall be limited to duration of initial license issued by AERB or the product liability period, whichever is longer.

- b. The term "supplier" and the duration and extent of supplier's liability are explained in Rule 24 of the Civil Liability for Nuclear Damage (CLND) Rules, 2011. For any questions relating to supplier's liability under section 17 of the Civil Liability for Nuclear Damage (CLND) Act, 2010, Government of India's clarifications dated February 08, 2015 may be referred to. These have been posted at the websites of Ministry of External Affairs and the Department of Atomic Energy under the title "Frequently Asked Questions and Answers on Civil Liability for Nuclear Damage Act 2010 and Related Issues".
- c. In regard to contracts with manufacturers of or vendors for supply of systems, equipment, components, or building of structures, or provision of services to nuclear installations which are operating or are under construction or those to be installed in future for which NPCIL is the system designer and technology owner, being responsible for safety design of such installations, NPCIL shall assume the role of supplier in accordance with the explanation of the term "supplier" given in Rule 24 of the CLND Rules, 2011 and in the context of section 17(a) and (b) of the CLND Act, 2010.
- d. Other suppliers can avail the Nuclear Suppliers' Special Contingency (Against Right of Recourse) Insurance Policy provided by the India Nuclear Insurance Pool to cover any liability exposure under section 17(a) and (b) of the CLND Act, 2010. NPCIL maintains the operator's statutory insurance under the CL

ND Act, 2010 by subscribing to the Nuclear Operator's Liability Policy offered by the India Nuclear Insurance pool, thereby subrogating to the India Nuclear Insurance pool the operator's "right of recourse" against suppliers under section 17(a) & (b) of the CLND Act, 2010.

- e. To have clarity on the terms used in the CLND Act 2010 and Rule 2011 pertaining to Right to Recourse, the following definition to be considered by the bidder before submission of bids.
- i. "Contractor" - shall be as per applicable GTC.
 - ii. "Supplier" shall be as defined in CLND Rule 24-2.
 - iii. "Product Liability Period (PLP)" shall be as defined in CLND Rule 24-2.
 - iv. "Initial License" (Refer CLND Rule 24-2): The initial license, unless otherwise specified, is valid for a period of five years from the date of its issue by AERB.
- f. **Right of Recourse under Civil Liability for Nuclear Damages Act 2010 & Rule 2011 thereof Not e:** Since requirement is for PHWR, NPCIL is the system designer and technology owner, being responsible for safety design of such installations in the this tender, NPCIL shall assume the role of Supplier in accordance with the explanation of term "Supplier" given in Rule 24 of the CLND Rules, 2011 and in the context of section 17(a) and (b) of the CLND Act, 2010.

8) Submission of Documents to consignee and Paying authority

a. Seller shall ensure submission of following documents to the Consignee along with the supply for the timely processing of Provisional Receipt Certificate (PRC) on GeM portal.

- i. Copy of Original invoice clearly marked for 'Bill to' and 'Shipped to'
- ii. e-Way Bill (In line with e-way bill provision under GST Act)
- iii. Delivery challan
- iv. Packing list (If applicable)

b. The Seller has to submit following documents to the Paying Authority immediately after despatching material to the Consignee:

- i. Original invoice clearly marked for 'Bill to' and 'Shipped to',
- ii. E-invoice if applicable to Seller based on his turnover
- iii. Annexure-F (format enclosed),
- iv. Seller's Bank details,
- v. Any other payment document mentioned in the contract required for paying authority

9) Liquidated Damages: Liquidated Damages as stipulated under clause no. 15 (iii) of GeM GTC shall be applicable for the contract awarded against this Bid

b.

10) In the event of "Dis-qualification" of Bid on account of non-compliance to bid conditions and non-submission of necessary documents against the Bid requirement and/or during the clarification on GeM Portal; The Clarifications/Documents submitted during the representation time shall not be considered for evaluation.

11) Representation/clarification in bid document:

Representation window is available on GeM portal for all bidders during the Bidding period. Bidder(s) shall use available window for any "Bid representation" or "Bid clarification" for appropriate response from buyer on GeM portal only. Bid clarification or representation raised through any other media shall be treated unsolicited.

12) Submission of Contractual Securities /Bank Guarantee

In case of award of contract, successful bidder shall furnish Performance Security for 5 % of the Contract value in any mode indicated in BID within 15 days of placement of contract for the duration of 17 months inclusive of 2 months claim period.

- 13)** Bidder/Seller can submit the Bank Guarantee through NeSL in eBG form. Details shall be used for NeSL as below:

Beneficiary PAN	AAACN3154F
Beneficiary Name	Site Director, Kakrapar Gujarat Site Nuclear Power Corporation of India Limited
Date of Incorporation	03/09/1987
Beneficiary Email Id	Debuy9946@npcil.co.in
Contact No:	02626-230271
Legal Constitution of the Party	Company
Registered Office Address	16 th Floor, World Trade Centre-1, Cuffe Parade, Colaba, Mumbai
Registered Office Address Pin code	400 005
Communication Address	Nuclear Power Corporation of India Limited Kakrapar Gujarat Site, Post-Anumala. Pin-394651, Taluka: Vyara, District : Tapi, Gujarat

- 14)** When GST is claimed by the seller/ supplier in general, the certificate as per Annexure F should be submitted to the paying authority with their invoices as per format given below:

<u>ANNEXURE - F</u>
GeM contract No.: GEMC _____
Invoice No. & Date_____
<u>Declaration Form for GST</u>
The Supplier/Contractors while submitting their bill to the Paying Authority shall furnish the following certificates:
Certified that: (Please Tick all appropriate options)
(a) Additional Input Tax Credit under GST availed against invoices submitted here under is Rs. _____.

- (b) Certified that the goods and services on which GST has been charged are not exempted under the GST Act or the rules made there under and the amount charged on account of GST on these goods and services are not more than what is payable under the relevant act or the rules there under.
- (c) Certified that we have taken into account all input tax credits available under GST and have not loaded the same in the basic price while furnishing their bids.
- (d) Certified that in respect of amount of taxes claimed in the bill no claim is pending for refund/or is admissible for refund from any other agency and /or no other tax credit is available in respect of the same. In the event of getting refund in whole or in part of the element of GST claimed from Government, the same shall be passed on the benefit to the Purchaser by remitting the amount equivalent to the amount of refund obtained.
- (e) Certified that the GST charged herein the invoices has been/ shall be deposited within the due date and the Invoice details have been / shall be populated in GSTR1/ANX-1 of the GSTN portal facilitating Input Tax Credit to the Purchaser.
- (f) Certified that we have complied with the Anti-profiteering measure provisions under CGST/ SGST/UTGST Acts and passed on commensurate reduction of price to the purchaser.
- (g) Certified that a quarterly statement shall be submitted to NPCIL confirming the payment of GST invoiced on NPCIL along with copy of GST paid Challan.
- (h) Certified that in case of any change in the Invoice which has been accepted by the Purchaser in the IMS portal, the corresponding amendment in the Original Invoice shall be made by the Supplier/Contractor, by raising Debit note/Credit note or any other document, as applicable, within the timelines stipulated under the GST Act.

Declaration Form for TDS

The Supplier/Contractors while submitting their bill to the Paying Authority shall furnish the following certificates:

Certified that: (Please Tick all appropriate boxes):

- (a) Whether Provisions of Section 194Q are applicable(For Supply of Goods): YES/NO
- (b) We have filed the return of income for both the financial years (F.Y.____ & F.Y.____)preceding the year in which tax is liable to be deducted within the prescribed time limit and hence, confirm that applicable rate of TDS is at Normal Rate (0.1% as on date)
- (c) We have linked the PAN to Aadhar and the PAN is not inoperative (applicable in case of Individuals)

Incomplete information against point (a), (b) & (c) above under Declaration form for TD S, shall be subject to deduction of tax at higher rate as applicable (5% as on date)

Declaration Form for Status of the Company

The Supplier/Contractors while submitting their bill to the Paying Authority shall furnish the following certificates:

Certified that: (Please Tick all appropriate boxes):

- (a) Whether Supplier is Company/LLP: YES/NO
- (b) Whether Company/LLP is declared as Strike Off Company/LLP in FY :Yes/NO

We have attached necessary documents downloaded from Ministry of Corporate Affairs

Signature of Contractor or their Authorised Representative with company seal.

IRN : VII098

15) Banning Of Business by NPCIL/BUYER

1.0	Banning of business dealings by NPCIL/Buyer
	<p>NPCIL reserves the right to initiate Banning as per NPCIL's Banning of business dealings as mentioned below and are independent of actions under GeM's IM (Incident Management) Policy.</p> <p>The words banning, blacklisting, de-registered, debarred, holiday, suspension of business etc., means the same.</p> <p>The words NPCIL, Corporation, Buyer etc., means the same.</p> <p>The words Contractor, Bidder, Seller, Seller, Service Provider etc., means the same.</p>
1.1	Grounds for Banning
	<p>The business dealing with the Contractor/Bidder/Seller/Service Provider shall be liable for banning, on account of the reasons attributable to them, which shall include, but not limited to the following:</p>
1.1.1.	Involvement in cartel formation during bidding.
1.1.2.	Baseless allegations by the bidder on NPCIL/Corporation/Buyer evaluation processes or officials.
1.1.3.	If any of the owner, proprietor or partner of the Contractor, is convicted by a court of law, during bidding process or currency of the contract, for offences involving corrupt and fraudulent practices including misrepresentation of the facts, moral turpitude in relation to its business dealings with NPCIL.
1.1.4.	Malafide / unlawful acts / malpractices or improper conduct on part of Contractor based on the approved findings of the Investigation Agency.
1.1.5.	If the Contractor misuses the premises or facilities of the NPCIL forcefully o

occupies, tampers or damages the Company's properties etc. or fails to vacate the properties/land/complex within reasonable time limit as specified or even after receiving the notices from the department.

1.1.6. Security concerns for the assets of the Corporation and State.

1.1.7. Submission of bids that contain false information or falsified documents or the concealment of such information in the bids in order to influence the outcome of eligibility screening or / at any other stage of the public bidding and execution.

1.1.8. Withdrawal of a bid or refusal to accept an award of contract with the NPCIL without justifiable cause, after being adjudged as the successful bidder.

1.1.9. Supply of Counterfeit items Breach of Code of Integrity.

1.1.10. Bidder shall not act in contravention of the codes which includes

1.1.10.1 Prohibition of

- a.** Making offer, solicitation or acceptance of bribe, reward or gift or any material benefit, either directly or indirectly, in exchange for an unfair advantage in the procurement process or to otherwise influence the procurement process.
- b.** Any omission or misrepresentation that may mislead or attempt to mislead so that financial or other benefit may be obtained or an obligation avoided.
- c.** Any conclusion, bid rigging or anticompetitive behavior that may impair the transparency, fairness and the progress of the procurement process.
- d.** Improper use of information provided by the procuring entity to the bidder with an intent to gain unfair advantage in the procurement process or for personal gain.
- e.** Any financial or business transactions between the bidder and any official of the procuring entity related to tender or execution process of contract; which can affect the decision of the procuring entity directly or indirectly.
- f.** Any coercion or any threat to impair or harm, directly or indirectly, any party or its property to influence the procurement process.
- g.** Obstruction of any investigation or auditing of a procurement process.
- h.** Making false declaration or providing false information for participation in tender process or to secure a contract;

1.1.10.2 Disclosure of conflict of interest

1.1.10.3 Disclosure by the bidder of any previous transgression made in respect of the provisions of above **1.1.10.1** with any entity in any country during the last three years or of being debarred/ banned by any other procuring entity.

1.2 Show Cause Notice

1.2.1. NPCIL will issue Show Cause Notice to the Contractor on noticing/receipt of a complaint of any irregularities and /or misconduct and /or unethical practice as mentioned in clause no. 1.1.

1.2.2. Upon receipt of Show cause notice, the Contractor is required to submit the reply to Show Cause Notice within 30 days of its receipt and no extension shall be given without justifiable reasons. The Contractor shall also be given an opportunity for oral hearing to present the case in person to NPCIL and the date of Oral Hearing will be indicated in the Show Cause Notice. Only the regular employees of Contractor or will be permitted to represent the Contractor during the Oral hearing, and no outsider shall be allowed to represent the Contractor on their behalf.

1.3 Period of Banning

The period of banning shall be for a period of not exceeding 2 (two) years and not less than 6 (six) months as considered appropriate by NPCIL.

1.4 Effect of Banning of Business Dealings by NPCIL

In case NPCIL has banned the business dealing with the bidder/contractor, the following shall be the consequences on issuance of the order of banning of business dealings with the bidder/contractor:

1.4.1 No Contract of any kind whatsoever shall be placed to such banned firms including its allied firms after the issue of Banning Order by NPCIL. The Contractor including their allied firms shall not be allowed to participate in any tender enquiry till completion of Banning period. If the Contractor has already participated in tender process and the price bids are not opened, his techno-commercial bid will be rejected and price bid will be kept unopened. In cases, where the price bids of Contractor have been opened prior to the order of banning, such bids shall be rejected. However, in case such banned Contractor is Lowest (L1), next lowest firm shall be considered as L1. Bid Security, if any, submitted by such banned Contractors shall be returned to the bidder.

1.4.2 Contractors shall not be permitted to participate in any business process in any form or entity i.e., as an Associate/Collaborator/Joint Venture Partner/Consortium Partner of the Main Contractor even if the banning order is passed subsequent to opening of Part-I bids.

1.4.3 Contractor shall not be allowed to participate as Sub-Vendor/sub-contractor in the tenders for contracts for works, service, supplies.

1.4.4 Even if, the banned Contractor is an approved Sub-vendor under any Contract for such equipment/component/service, the Main Contractor shall not be permitted to place Work order/Purchase order/Service Contract on the banned Contractor as a sub-vendor after the date of banning even though the name of the party has been approved as a sub-vendor prior to the order of banning.

1.4.5 The completion certificate issued to the contractor shall make a mention regarding banning during execution of the contract.

1.4.6 Banned bidders shall not be permitted to submit their bid. The bid submitted by the banned bidder shall be summarily rejected.

1.4.7 Contracts concluded before the issue of the banning order shall, not be affected by the banning order.

1.4.8 Banning shall automatically be extended to all Allied firms of the Contractor. In case of Joint venture/ Consortium is banned all partners will also stand debarred for the period specified in the Banning Order. The names of all partners should be clearly specified in the "Banning order".

1.4.9 Banning in any manner does not impact any other contractual or other legal rights of NPCIL.

1.4.10 Banning under the provisions of Banning of Business Dealings of NPCIL is applicable only for NPCIL.

1.5 Definition of Allied Firm:

Allied Firm means all concerns which come within the sphere of effective influence of the banned firm. In determining this, the following factors shall be taken into consideration:

- a.** Whether the management is common;
- b.** Majority interest in the management is held by the partners or directors of banned/

- c.** suspended firm;
- d.** Substantial or majority shares are owned by the banned/ suspended firm and by virtue
- e.** of this it has a controlling voice;
- f.** Directly or indirectly controls, or is controlled by or is under common control with another bidder;
- g.** All successor firms will also be considered allied firms.

16) Submission/Exemption of EMD shall be as per GeM GTC. However, Vendor Assessment is not applicable for Custom/BOQ bids. Hence, "Vendor Assessment Report" shall not be considered for EMD exemption in the instant bid. Accordingly, EMD exemption shall be allowed as per GeM GTC, except against "Vendor Assessment Report".

17) Bidder shall provide Compliance to Annexure II attached in the Bid Document.

Annexure-II

(Bidder is required to submit following Undertaking on Company's Letter Head)

घोषणापत्र/UNDERTAKING

मैं एतद्वारा वचन देता हूँ कि, जीईएम जीटीसी खंड संख्या -29 (प्रति बिडर एक बिड) के अनुपालन में, मैंने इस निविदा में केवल एक बोली प्रस्तुत की है। जीईएम जीटीसी क्लॉज नंबर -29 का उल्लंघन, इस निविदा के मूल्यांकन या बाद में प्रसंस्करण के दौरान या कार्य के निष्पादन के दौरान मनाया जाता है, जिसके परिणामस्वरूप बोली (ओं) की अयोग्यता/अनुबंध रद्दकरण, ईएमडी की जब्ती और फर्म के खिलाफ आवश्यक प्रशासनिक कार्रवाई शुरू हो जाएगी। जीईएम दिशानिर्देशों के अनुसार। मैं ऐसे सभी प्रशासनिक कार्यों के लिए पूरी तरह से जिम्मेदार रहूंगा।

I hereby undertake that, in compliance to GeM GTC Clause No.-29 (One Bid per Bidder), I have submitted only one bid in this tender. Violation of GeM GTC Clause No.-29, observed during evaluation or subsequent processing of this tender or during execution of work, will result in bid(s) disqualification/contract cancellation, forfeiture of EMD and initiation of necessary administrative actions against the firm (s) as per GeM guidelines. I will be solely responsible for all such administrative actions.

यह प्रमाणित किया जाता है कि/It is certified that:

- विषय निविदा में केवल एक बोली प्रस्तुत की गई है।/Only one bid has been submitted in the subject tender.
- किसी भी सहयोगी/संबद्ध संस्थाओं ने विषय निविदा में भाग नहीं लिया है।/Any sister/Associated/Allied concerns have not participated in the subject tender.
- सामान्य व्यवसाय स्वामित्व/प्रबंधन वाली एक से अधिक इकाइयों के मामले में निविदा में केवल एक इकाई ने भाग लिया है।/ Only one unit has participated in the tender, in case of more than one unit having common business ownership/management.
- जीईएम जीटीसी क्लॉज नंबर -29 का पूरी तरह से अनुपालन किया गया है।/ GeM GTC Clause No.-29 is fully complied.

विधिवत हस्ताक्षरित और मुहरबंद/ Du
ly signed & sealed by:

फर्म/कंपनी के अधिकृत प्रतिनिधि/ (Authorised Representative of the firm/company)

नाम और पद धारित आयोजित/Name
& Post held:

दिनांक/Dated:

12. **Forms of EMD and PBG**

Bidders can also submit the EMD with Account Payee Demand Draft in favour of

NPCIL, KAPS VYARA
payable at
NPCIL, KAPS VYARA

Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

13. **Forms of EMD and PBG**

Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C

Nuclear Power Corporation of India Limited, VYARA

The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date/ Bid Opening date

14. **Forms of EMD and PBG**

Bidders can also submit the EMD with Banker's Cheque in favour of

NPCIL, KAPS VYARA
payable at
NPCIL, KAPS VYARA

Bidder has to upload scanned copy / proof of the BC along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

15. **Forms of EMD and PBG**

Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name

Nuclear Power Corporation of India Limited
Account No.
10394473111
IFSC Code
SBIN0000532
Bank Name
State Bank of India
Branch address
Vyara

Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.

16. **Forms of EMD and PBG**

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of

NPCIL, KAPS VYARA
payable at
NPCIL, KAPS VYARA

After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

17. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of

Nuclear Power Corporation of India Limited, VYARA

A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.

18. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Payment online through RTGS / internet banking also (besides PBG which is allowed as per GeM GTC). On-line payment shall be in Beneficiary name

Nuclear Power Corporation of India Limited

Account No.

10394473111

IFSC Code

SBIN0000532

Bank Name

State Bank of India

Branch address

Vyara

. Successful Bidder to indicate Contract number and name of Seller entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer in place of PBG within 15 days of award of contract.

अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.

12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---